



Continuation of Substance of Interview including description of the general nature of what was discussed: The examiner contacted applicant's representative to request permission to make an examiner's amendment to remove the intended use issues from applicant's apparatus claims, correct numerous grammatical errors, and clarify that applicant's claims are drawn to a "mutinuclear frequency adjustable detachable wand" which is fully supported in applicant's original disclosure, and originally submitted title, as opposed to the prior art which detects different nuclei by switching between a series of individually frequency tuned capacitor sticks. In applicant's invention the slidable collar enables different nuclei to be detected without requiring a separate capacitor stick. The examiner also informed applicant's representative that the submitted replacement drawings of 08/13/2004 have stray toner lines from faxing and that a complete clean complete set of formal drawings would need to be submitted. The examiner was thanked for her time, and authorization to perform the examiner's amendment was given in order to place the application in condition for allowance. .